



1600 E. Murdock Wichita, KS 67214  
Phone: (866) 384-3010



**Pressure Gauge**  
Calibration Report for:

**NRG Energy - High Lonesome Mesa**  
550 Kenshaw Rd  
Willard, NM 87063

Calibration #:	150512.26.48369
Calibration Date:	05/12/2015
Expiration Date:	05/12/2016
Temperature:	23.7 °C
Humidity:	11.70% RH

Asset ID #:	1431454901
Asset Type:	Pressure Gauge
Manufacturer:	ITH
Model #:	EN 837-1
Serial #:	082753033

Set Point (BAR)	"As Found" Data (1) (BAR)	"As Found" CL 95% (BAR)	"As Left" Data (2) (BAR)	"As Left" CL 95% (BAR)	Low Limit -2% (BAR)	High Limit +2% (BAR)
160	179.2	.0694	179.2	.0694	128	192
320	337.8	0.2	337.8	0.2	288	352
640	655.1	0.3	655.1	0.3	608	672
960	969.6	2.9	969.6	2.9	928	992
1280	1289.2	2.9	1289.2	2.9	1248	1312
1600	1610.9	2.9	1610.9	2.9	1568	1632
1280	1277.1	2.9	1277.1	2.9	1248	1312
960	956.5	2.9	956.5	2.9	928	992
640	645.2	0.3	645.2	0.3	608	672
320	328.5	0.2	328.5	0.2	288	352
160	168.8	.069	168.8	.069	128	192

**Remarks:**

- (1) "As Found" Data: **IN TOLERANCE**  
Test Instrument was found within applicable tolerance levels.
- (2) "As Left" Data: **IN TOLERANCE**  
Test Instrument was left within applicable tolerance levels.

**Technician Comments:**

Jason Hankins, Calibration Technician

**Notes:**

- A. This accredited organization maintains both A2LA accreditation to ISO/IEC 17025 and ANSI/NCSL Z540-1-1994 for the specific tests/calibrations listed in A2LA Certificate #3040.01, and ISO 6789.
- B. Environmental conditions were held within Alltite's standards of 60 to 88 degrees Fahrenheit and below 90% relative humidity.
- C. Procedures for this calibration method are outlined in CP-006-19.
- D. All standards and associated equipment used to calibrate this instrument have been calibrated and are traceable to A2LA accreditation and the National Institute of Standards and Technology (NIST).
- E. Customer acknowledges and agrees that Alltite's standard sales terms and conditions are incorporated herein by reference and made a part of this calibration report.
- F. This document relates only to the test instrument listed on this certificate and cannot be reproduced, except in full, without written approval of Alltite.
- G. This calibration document is followed by a Calibration Schedule page.
- H. The uncertainty(s) listed represent an expanded uncertainty expressed using a coverage factor of k=2, such that the confidence level (CL) approximates 95%. CL 95% = Unaccred. or \* means the measurement falls outside our accredited scope and therefore no measure of uncertainty has been determined.



## Van 04 : WKSHQ 7629 (Inactive) Calibration Laboratory: Certification Schedule

Capacity	Item	Model #	Serial #	Traceability/Report #	Date Due Cal
36,000 psi	Additel	681	211H13490010	46660.001	11/15/2015
Temp/Humid	Extech (gave to Lockhart)	42280	10075032	1001667023	01/08/2016