

REPORT OF CALIBRATION

Alltite, Inc.
1600 E. Murdock
Wichita, KS, 67214

UNIT UNDER TEST:	Fluke PRV240 Proving Unit	TEST RESULTS:	In Tolerance
SERIAL NUMBER:	47851156	CAL DATE:	18-Jun-2021
CID NUMBER:	210618.58.31144	CAL DUE:	18-Jun-2022
PROCEDURE NAME:	Report of Calibration PRV240 (Unaccred)	DATA TYPE:	AS FOUND/AS LEFT
PROCEDURE REVISION:	1.0	TEMPERATURE:	21.4 °C
P.O. NUMBER:		HUMIDITY:	53.0 %
CALIBRATED BY:	M. Coker		
CUSTOMER:	Avangrid Renewables -Otter Creek		
ADDRESS1:	2706 N 13 th Rd		
City, State:	Ransom, IL 60470		

Alltite, INC. certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure unless otherwise noted. This Report of Calibration applies only to the item being calibrated, identified above. The report may contain data that is not within the Scope of Accreditation. The unaccredited material is either confined to sections of this certificate that are clearly marked or as "Not Accredited" or is indicated by the lack of a displayed value of "Measurement Uncertainty" adjacent to affected parameter or is indicated by the word "No" in the Accred. column.

This calibration report documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI). The user may be obliged to have the object recalibrated at appropriate intervals. This report applies to only the item identified and shall not be reproduced, except in full, unless written permission for an approved abstract is obtained from Alltite, INC. This report shall not be used to claim product endorsement by the accrediting body. Calibration reports without Authorizing Signatures are not valid.

This report makes no claim of compliance or non-compliance. However, measured values greater than the specification limits are identified in the report by ">Spec." Measured values are expressed in terms of the indicated parameter, for example, mV, A, etc.

Measurement uncertainties at the time of test are expressed in the base units, are given on the following pages, where applicable. They are calculated in accordance with the methods described in EA-4/02, NIST TN1297, DKD-3, or applicable Guides to the Uncertainty Measurement (GUM), using a coverage factor of $k=2$, corresponding to a confidence level of approximately 95%.

This calibration was performed using measurement standards traceable to the applicable National Measurement Institute (NMI), to accepted intrinsic standards of measurement, or is derived by ratio type self-calibration techniques. The calibration system complies with the requirements of NIST Handbook 150, ANSI.NCSL Z540.1-1999, ISO/IEC 17025, EN45001 and Guidelines DKD 1 to 7. $CMC^2 \pm = 0.000$ or not displayed, the reading is **not** an accredited calibration point.

Remarks:

As Found:
As Left:

Standards Used

Serial Number	Manufacturer / Model	Cal Date	Cal Due
4554802	FLUKE 5502A	19-Aug-2020	19-Feb-2022

Test Results

Step	Range	Calibrator Output	Pass / Fail
1	240VDC $\pm 10\%$	239.086	
2	240VAC $\pm 10\%$	240.651	

**** End of Certificate ****